



## Certificate of Analysis

**Product name**

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

**Material No.**

05278368001

**Lot No.**

Y18581

**Date of Manufacturing**

2017-09-29

**Expiry Date**

2019-03-15

**Parameter****Result****Unit**

Functional Staining

Pass

Breast Carcinoma Tissue, Gastric Carcinoma Tissue,  
and 4 in 1 Cell Line.The specific staining of the test batch must be  
acceptable on all tissues and cell line cores using  
both the iView DAB IHC and the ultraView DAB IHC  
detection kits.The test batch background intensity must not exceed  
0.5 pt on any tissue or cell line core.All tissues and cell line cores stained with the negative  
control reagent must be negative for specific staining.

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and  
substantially free of intrinsic particulate matter**Issue date** 2017-10-11

Page 1 of 2



## Certificate of Analysis

**Product name**

PATHWAY ANTI-HER-2/NEU (4B5) RABBIT MONO

**Material No.**

05278368001

**Lot No.**

Y18581

**Date of Manufacturing**

2017-09-29

**Expiry Date**

2019-03-15

**Parameter****Result****Unit**

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

**Disposition by**

Whitney Croft

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2017-10-11

Page 2 of 2

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 520 887 2155  
<http://www.ventana.com>