



# VENTANA anti-p63 (4A4) Mouse Monoclonal Primary Antibody

REF

790-4509

05867061001

IVD





Figure 1. VENTANA anti-p63 (4A4) antibody staining of normal basal cells in prostate tissue.

# INTENDED USE

VENTANA anti-p63 (4A4) Mouse Monoclonal Primary Antibody is intended for laboratory use in the qualitative immunohistochemical detection of the p63 protein by light microscopy in sections of formalin-fixed, paraffin-embedded tissue stained on a BenchMark IHC/ISH instrument.

This product should be interpreted by a qualified pathologist in conjunction with histological examination, relevant clinical information, and proper controls.

This antibody is intended for in vitro diagnostic (IVD) use.

# **SUMMARY AND EXPLANATION**

VENTANA anti-p63 (4A4) Mouse Monoclonal Primary Antibody (VENTANA anti-p63 (4A4) antibody) is a mouse monoclonal antibody produced against the p63 protein. The human tumor protein 63 (TP63, p63) is a 77 kDa protein localized to the cellular nucleus and member of the p53 family of transcription factors. The p63 protein is expressed in the basal or progenitor cells of epithelial tissues and glandular structures including the prostate, breast, and bronchi. 2,3,4 In the prostate, p63 is expressed in the basal cells of almost all normal and benign glands but is not present in the neuroendocrine or luminal secretory cells. In breast, p63 is expressed in the basal myoepithelial cell (MEC) layer in normal tissue as well as in benign lesions. Invasive lesions that evolve from and involve the prostate or breast can disrupt and eventually breach the basal membrane thus eliminating the presence of prostatic basal cells or breast MECs. 1,6

The immunohistochemistry (IHC)-based detection of p63 in prostatic basal cells is a feature of normal and benign processes and the absence of p63 is indicative of carcinoma of the prostate. Fig. 10 the detection of p63 in prostatic basal cells with VENTANA anti-p63 (4A4) antibody may be used to aid in the differentiation of benign and malignant prostate lesions. The detection of p63 in breast MECs is a hallmark of non-invasive processes and the absence of p63 is indicative of invasive neoplasms. Fig. 10 the detection of p63 in breast MECs with VENTANA anti-p63 (4A4) antibody may be used to aid in distinguishing invasive from non-invasive breast neoplasms.

In lung, p63 is expressed in the basal cell compartment.<sup>3</sup> It has been speculated that squamous cell carcinoma (SCCA) of the lung originates from the basal compartment.<sup>9</sup> Thus, the overexpression of p63 in non-small cell lung cancer (NSCLC) can be an indicator of malignant squamous differentiation. <sup>10,11,12</sup> The detection of p63 with VENTANA anti-p63 (4A4) antibody may be used as a marker of squamous differentiation to aid in the distinction between pulmonary SCCA and pulmonary adenocarcinoma (ADC). This antibody may be used as part of a panel of IHC studies.

## PRINCIPLE OF THE PROCEDURE

VENTANA anti-p63 (4A4) antibody is a mouse monoclonal antibody produced against a recombinant fragment of the N-terminal sequence of the human p63 protein. VENTANA anti-p63 (4A4) antibody binds to the p63 protein in formalin-fixed, paraffin-embedded (FFPE) tissue sections and exhibits a nuclear staining pattern. This antibody can be visualized using *ultra*View Universal DAB Detection Kit (Cat. No. 760-500 / 05269806001). Refer to the respective method sheets for further information.

## **MATERIAL PROVIDED**

VENTANA anti-p63 (4A4) antibody contains sufficient reagent for 50 tests.

One 5 mL dispenser of VENTANA anti-p63 (4A4) antibody contains approximately 0.7  $\mu g$  of a mouse monoclonal antibody.

The antibody is diluted in Tris-HCl with carrier protein and 0.10% ProClin 300, a preservative.

Specific antibody concentration is approximately  $0.14~\mu g/mL$ . There is no known non-specific antibody reactivity observed in this product.

VENTANA anti-p63 (4A4) is a monoclonal antibody produced as cell culture supernatant material

Refer to the appropriate VENTANA detection kit method sheets for detailed descriptions of: Principle of the Procedure, Material and Methods, Specimen Collection and Preparation for Analysis, Quality Control Procedures, Troubleshooting, Interpretation of Results. and Limitations.

## MATERIALS REQUIRED BUT NOT PROVIDED

Staining reagents, such as VENTANA detection kits and ancillary components, including negative and positive tissue control slides, are not provided.

Not all products listed in the method sheet may be available in all geographies. Consult your local support representative.

The following reagents and materials may be required for staining but are not provided:

- 1. Recommended control tissue
- 2. Microscope slides, positively charged
- 3. Negative Control (Monoclonal) (Cat. No. 760-2014 / 05266670001)
- 4. ultraView Universal DAB Detection Kit (Cat. No. 760-500 / 05269806001)
- 5. EZ Prep Concentrate (10X) (Cat. No. 950-102 / 05279771001)
- 6. Reaction Buffer Concentrate (10X) (Cat. No. 950-300 / 05353955001)
- 7. LCS (Predilute) (Cat. No. 650-010 / 05264839001)
- 8. ULTRA LCS (Predilute) (Cat. No. 650-210 / 05424534001)
- 9. Cell Conditioning Solution (CC1) (Cat. No. 950-124 / 05279801001)
- 10. ULTRA Cell Conditioning Solution (ULTRA CC1) (Cat. No. 950-224 / 05424569001)
- 11. Hematoxylin II (Cat. No. 790-2208 / 05277965001)
- 12. Bluing Reagent (Cat. No. 760-2037 / 05266769001)
- 13. Permanent mounting medium
- 14. Cover glass
- 15. Automated coverslipper
- 16. General purpose laboratory equipment
- 17. BenchMark IHC/ISH Instrument

# STORAGE AND STABILITY

Upon receipt and when not in use, store at 2-8°C. Do not freeze.

To ensure proper reagent delivery and the stability of the antibody, replace the dispenser cap after every use and immediately place the dispenser in the refrigerator in an upright position.

Every antibody dispenser is expiration dated. When properly stored, the reagent is stable to the date indicated on the label. Do not use reagent beyond the expiration date.

# **SPECIMEN PREPARATION**

Routinely processed FFPE tissues are suitable for use with this primary antibody when used with VENTANA detection kits and BenchMark IHC/ISH instruments. The recommended tissue fixative is 10% neutral buffered formalin.  $^{13}$  Sections should be cut at approximately 4  $\mu m$  in thickness and mounted on positively charged slides. Slides should be stained immediately, as antigenicity of cut tissue sections may diminish over time. Ask your Roche representative for a copy of "Recommended Slide Storage and Handling" for more information.

It is recommended that positive and negative controls be run simultaneously with unknown specimens.





## WARNINGS AND PRECAUTIONS

- 1. For in vitro diagnostic (IVD) use.
- 2. For professional use only.
- 3. Do not use beyond the specified number of tests.
- 4. ProClin 300 solution is used as a preservative in this reagent. It is classified as an irritant and may cause sensitization through skin contact. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Use protective clothing and gloves.
- Positively charged slides may be susceptible to environmental stresses resulting in inappropriate staining. Ask your Roche representative for more information on how to use these types of slides.
- Materials of human or animal origin should be handled as biohazardous materials and disposed of with proper precautions. In the event of exposure, the health directives of the responsible authorities should be followed.<sup>14,15</sup>
- Avoid contact of reagents with eyes and mucous membranes. If reagents come in contact with sensitive areas, wash with copious amounts of water.
- 8. Avoid microbial contamination of reagents as it may cause incorrect results.
- For further information on the use of this device, refer to the BenchMark IHC/ISH instrument User Guide, and the instructions for use of all necessary components located at navifyportal.roche.com.
- Consult local and/or state authorities with regard to recommended method of disposal.
- Product safety labeling primarily follows EU GHS guidance. Safety data sheets available on request.
- To report suspected serious incidents related to this device, contact the local Roche representative and the competent authority of the Member State or Country in which the user is established.

This antibody or assay contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

Table 1. Hazard Information.

Hazard	Code	Statement
Warning	ng H317 May cause an allergic skin reaction.	
	H412	Harmful to aquatic life with long lasting effects.
	P261	Avoid breathing mist or vapours.
	P273	Avoid release to the environment.
	P280	Wear protective gloves.
P333 + If skin irritation or attention.		If skin irritation or rash occurs: Get medical advice/ attention.
	P362 + P364	Take off contaminated clothing and wash it before reuse.
	P501	Dispose of contents/ container to an approved waste disposal plant.

This product contains CAS # 55965-84-9, reaction mass of: 5-chloro-2-methyl-2H-isothiazol-3-one and 2-methyl-2H-isothiazol-3-one (3:1).

# STAINING PROCEDURE

VENTANA primary antibodies have been developed for use on BenchMark IHC/ISH instruments in combination with VENTANA detection kits and accessories. Refer to Table 2 for recommended staining protocols.

This antibody has been optimized for specific incubation times but the user must validate results obtained with this reagent.

The parameters for the automated procedures can be displayed, printed and edited according to the procedure in the instrument User Guide. Refer to the appropriate VENTANA detection kit method sheet for more details regarding immunohistochemistry staining procedures.

For more details on the proper use of this device, refer to the inline dispenser method sheet associated with P/N 790-4509.

**Table 2.** Recommended staining protocol for VENTANA anti-p63 (4A4) antibody with *ultra*View Universal DAB Detection Kit on BenchMark IHC/ISH instruments.

	Method		
Procedure Type	ХТ	ULTRA or ULTRA PLUS <sup>a</sup>	
Deparaffinization	Selected	Selected	
Cell Conditioning (Antigen Unmasking)	Cell Conditioning 1, Standard	ULTRA Cell Conditioning 1 Standard	
Antibody (Primary)	16 minutes, 37 °C	20 minutes, 36 °C	
Counterstain	Hematoxylin II, 4 minutes		
Post Counterstain	Bluing, 4 minutes		

<sup>&</sup>lt;sup>a</sup> Concordance was demonstrated between BenchMark ULTRA and BenchMark ULTRA PLUS instruments using representative assays.

Due to variation in tissue fixation and processing, as well as general lab instrument and environmental conditions, it may be necessary to increase or decrease the primary antibody incubation, cell conditioning or protease pretreatment based on individual specimens, detection used, and reader preference. For further information on fixation variables, refer to "Immunohistochemistry Principles and Advances." 16

## **NEGATIVE REAGENT CONTROL**

In addition to staining with VENTANA anti-p63 (4A4) antibody, a second slide should be stained with the appropriate negative control reagent.

#### POSITIVE TISSUE CONTROL

Optimal laboratory practice is to include a positive control section on the same slide as the test tissue. This helps identify any failures applying reagents to the slide. Tissue with weak positive staining is best suited for quality control. Control tissue may contain both positive and negative staining elements and serve as both the positive and negative control. Control tissue should be fresh autopsy, biopsy, or surgical specimen, prepared or fixed as soon as possible in a manner identical to test sections.

Known positive tissue controls should be utilized only for monitoring performance of reagents and instruments, not as an aid in determining specific diagnosis of test samples. If the positive tissue controls fail to demonstrate positive staining, results of the test specimen should be considered invalid.

Examples of positive control tissues for this antibody are normal prostate and normal tonsil

## STAINING INTERPRETATION / EXPECTED RESULTS

The cellular staining pattern for VENTANA anti-p63 (4A4) antibody is nuclear.

# **SPECIFIC LIMITATIONS**

This antibody has been optimized for incubation time on a BenchMark IHC/ISH instrument, in combination with *ultra*View Universal DAB Detection Kit, but the user must validate results obtained with this reagent.

All assays might not be registered on every instrument. Please contact your local Roche representative for more information.

## PERFORMANCE CHARACTERISTICS

## **ANALYTICAL PERFORMANCE**

Staining tests for specificity, sensitivity, and repeatability were conducted and the results are listed below.





# **Sensitivity and Specificity**

**Table 3.** Sensitivity/Specificity of VENTANA anti-p63 (4A4) antibody was determined by testing FFPE normal tissues.

Tissue	# positive / total cases	Tissue	# positive / total cases
Cerebrum	0/3	Thymus	2/3
Cerebellum	0/3	Myeloid (Bone marrow)	0/3
Adrenal gland	0/3	Mesothelium and lung	0/13
Ovary	0/3	Heart	0/3
Pancreas	0/3	Esophagus	1/3
Parathyroid gland	0/3	Stomach	0/3
Hypophysis (Pituitary)	0/3	Small intestine	0/3
Testis <sup>a</sup>	2/3	Colon	0/3
Thyroid	0/3	Liver	0/3
Breast	21/21	Salivary gland <sup>a</sup>	3/3
Spleen	0/3	Kidney	0/3
Tonsil	3/3	Prostate	64/65
Endometrium	0/3	Cervix	3/3
Skeletal muscle	0/3	Skin	3/3
Peripheral Nerve	0/3		-

a Focal staining

**Table 4.** Sensitivity/Specificity of VENTANA anti-p63 (4A4) antibody was determined by testing a variety of FFPE neoplastic tissues.

Pathology	# positive / total cases
Glioblastoma (Cerebrum) a	1/1
Meningioma (Cerebrum)	1/1
Ependymoma (Cerebrum)	0/1
Oligodendroglioma (Cerebrum)	0/1
Serous carcinoma (Ovary)	0/1
Mucinous carcinoma (Ovary)	0/1
Neuroendocrine neoplasm (Pancreas)	0/1
Adenocarcinoma (Pancreas)	1/1
Seminoma (Testis)	0/1
Embryonal carcinoma (Testis)	0/1
Medullary carcinoma (Thyroid)	0/1
Papillary carcinoma (Thyroid)	0/1
B-cell lymphoma, NOS (Spleen)	0/1
Small cell carcinoma (Lung)	0/1
Squamous cell carcinoma (Lung) b	131/150
Adenocarcinoma (Lung) <sup>b</sup>	12/116
Squamous cell carcinoma (Esophagus)	1/1
Adenocarcinoma (Esophagus)	0/1
Mucinous adenocarcinoma (Stomach)	0/1
Adenocarcinoma (Gastrointestinal)	0/3

Pathology	# positive / total cases
Gastrointestinal stromal tumor (GIST)	0/3
Hepatocellular carcinoma (Liver)	0/1
Hepatoblastoma (Liver)	0/1
Clear cell carcinoma (Kidney)	0/1
Adenocarcinoma (Prostate)	1/113
Leiomyosarcoma (Prostate)	0/1
Urothelial carcinoma (Prostatic urethra)	1/1
Leiomyoma (Uterus)	0/1
Carcinoma (Uterus)	0/1
Clear cell carcinoma (Uterus)	0/1
Squamous cell carcinoma (Uterus)	1/2
Embryonal rhabdomyosarcoma (Striated muscle) a	1/1
Melanoma (Anus)	0/1
Basal cell carcinoma (Skin)	1/1
Squamous cell carcinoma (Skin)	1/1
Neurofibroma (Lumbar)	0/1
Neuroblastoma (Retroperitoneum)	0/1
Epithelioid mesothelioma	0/1
Lymphoma, NOS (Lymph node)	2/3
Hodgkin lymphoma (Lymph node)	0/1
Urothelial carcinoma (Bladder)	1/1
Leiomyosarcoma (Bladder)	0/2
Osteosarcoma	0/1
Spindle cell rhabdomyosarcoma (Peritoneum)	0/1
Lobular carcinoma in situ (LCIS) (Breast) c	3/3
Ductal carcinoma in situ (DCIS) (Breast) c	8/10
Invasive ductal carcinoma (Breast)	2/9
Invasive lobular carcinoma (Breast)	0/8

a Focal staining.

# Precision

Precision studies for VENTANA anti-p63 (4A4) antibody were completed to demonstrate:

- Between lot precision of the antibody.
- Within run and between day precision on a BenchMark XT instrument.
- Between instrument precision on the BenchMark XT and BenchMark ULTRA instrument.
- Between platform precision between the BenchMark XT and BenchMark ULTRA instrument.

All studies met their acceptance criteria.

Precision on the BenchMark ULTRA PLUS instrument was demonstrated using representative assays. Studies included Within-run Repeatability, Between-day and Between-run Intermediate Precision. All studies met their acceptance criteria.

<sup>&</sup>lt;sup>b</sup> Some cases demonstrated focal staining.

<sup>&</sup>lt;sup>c</sup> Normal myoepithelial cells staining positive.





## **CLINICAL PERFORMANCE**

Clinical performance data relevant to the intended purpose of VENTANA anti-p63 (4A4) antibody were assessed by systematic review of the literature. The data gathered support the use of the device in accordance with its intended purpose.

## **REFERENCES**

- Yang A, Kaghad M, Wang Y, et al. p63, a p53 homolog at 3q27-29, encodes multiple products with transactivating, death-inducing, and dominant-negative activities. Mol Cell. 1998:2:305-16
- Yang A, Schweitzer R, Sun D, et al. p63 is essential for regenerative proliferation in limb, craniofacial and epithelial development. Nature. 1999;398:714-718.
- Nobre AR, Albergaria A, Schmitt F. P40: A P63 Isoform Useful for Lung Cancer Diagnosis - a Review of the Physiological and Pathological Role of P63. Acta Cytol. 2013:57(1):1-8.
- Barbareschi M, Pecciarini L, Cangi MG, et al. P63, a P53 Homologue, Is a Selective Nuclear Marker of Myoepithelial Cells of the Human Breast. Am J Surg Pathol. 2001;25(8):1054-1060.
- Signoretti S, Waltregny D, Dilks J, et al. p63 is a prostate basal cell marker and is required for prostate development. Am J Pathol. 2000;157:1769-1775.
- Epstein JI, Egevad L, Humphrey PA, et al. Best Practices Recommendations in the Application of Immunohistochemistry in the Prostate: Report from the International Society of Urologic Pathology Consensus Conference. Am J Surg Pathol. 2014;38(8):e6-e19.
- Zhao L, Yang X, Khan A, et al. Diagnostic Role of Immunohistochemistry in the Evaluation of Breast Pathology Specimens. Arch Pathol Lab Med. 2014;138(1):16-24
- Dewar R, Fadare O, Gilmore H, et al. Best Practices in Diagnostic Immunohistochemistry: Myoepithelial Markers in Breast Pathology. Arch Pathol Lab Med. 2011;135(4):422-429.
- 9. Hanna JM, Onaitis MW. Cell of Origin of Lung Cancer. J Carcinog. 2013;12:6.
- Warth A, Muley T, Herpel E, et al. Large-Scale Comparative Analyses of Immunomarkers for Diagnostic Subtyping of Non-Small-Cell Lung Cancer Biopsies. Histopathology. 2012;61(6):1017-1025.
- Rekhtman N, Ang DC, Sima CS, et al. Immunohistochemical Algorithm for Differentiation of Lung Adenocarcinoma and Squamous Cell Carcinoma Based on Large Series of Whole-Tissue Sections with Validation in Small Specimens. Mod Pathol. 2011;24(10):1348-1359.
- Tacha D, Yu C, Bremer R, et al. A 6-antibody panel for the classification of lung adenocarcinoma versus squamous cell carcinoma. Appl Immunohistochem Mol Morphol. 2012;20:201-207.
- Carson FL, Cappellano C. Histotechnology; A Self-Instructional Text, 5th edition. American Society for Clinical Pathology Press; 2020, 2022.
- Occupational Safety and Health Standards: Occupational exposure to hazardous chemicals in laboratories. (29 CFR Part 1910.1450). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 24 June 2020 on the protection of workers from risks related to exposure to biological agents at work.
- Roche PC, Hsi ED. Immunohistochemistry-Principles and Advances. Manual of Clinical Laboratory Immunology. 6th edition. In: NR Rose. ed. ASM Press: 2002.

**NOTE:** A point (period/stop) is always used in this document as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

The summary of safety and performance can be found here:

https://ec.europa.eu/tools/eudamed

## **Symbols**

Ventana uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see elabdoc.roche.com/symbols for more information).



Global Trade Item Number

Rx only

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

# **REVISION HISTORY**

Rev	Updates	
Н	Updates to Warnings and Precautions section. Updated to current template.	

## INTELLECTUAL PROPERTY

VENTANA, BENCHMARK, and ULTRAVIEW are trademarks of Roche. All other product names and trademarks are the property of their respective owners.

© 2024 Ventana Medical Systems, Inc.

For USA: Rx only

# **CONTACT INFORMATION**



Ventana Medical Systems, Inc. 1910 E. Innovation Park Drive Tucson, AZ 85755 USA

- +1 520 887 2155
- +1 800 227 2155 (USA)

www.roche.com



Roche Diagnostics GmbH Sandhofer Strasse 116 68305 Mannheim Germany

+800 5505 6606

