



Certificate of Analysis

Product name

VENTANA ANTI-P120 CATENIN (98)

Material No.

05867088001

Lot No.

K14788

Date of Manufacturing

2023-08-04

Expiry Date

2025-08-03

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Functional Staining

Pass

Ductal and Lobular Breast Carcinoma Tissues

The test batch and reference batch must score 2.0 or greater on both ductal and lobular breast carcinoma tissues for specific stain intensity on a 0 to 3 scale.

The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch.

The background intensity of the test batch, reference batch, and negative reagent control must be 0.5 or less on both tissues

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Mia Cazares

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2023-08-28

Page 1 of 1

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>