

CERTIFICATE OF ANALYSIS

Product Name: Iron Staining Kit
Ventana REF: 860-009
Roche Number: 05279259001
Batch Number: D11202

Test	Acceptance Criteria	Result
Functional Staining on liver and bone marrow tissue	Ferrous iron should stain bright blue against a background of pink to red	PASS
	In bone marrow immature erythrocytes may stain very lightly blue-green	PASS
	Negative Control – pink to red counterstained tissue	PASS


This lot met the specifications.

Expiry Date: 2016-01-28

Manufacture Date : 08-FEB-2014

Elizabeth Garcia

Quality Department

 12-Feb-2014

Date