



Certificate of Analysis

Product name
CONFIRM Anti-PR (1E2) Primary Antibody

Material No.
05277990001 or 05278392001

Lot No.	Date of Manufacturing	Expiry Date
E05750	23-Mar-2018	2020/03/11

Parameter	Result	Unit
------------------	---------------	-------------

Visual Inspection	Pass	
-------------------	------	--

Reagent must be free of foreign material and substantially free of particulate matter

Reagent must be clear and free of turbidity

Functional Staining	Pass	
---------------------	------	--

Breast Carcinoma Tissue

The nuclear staining will be interpreted according to the following approximate percent tumor cell staining criteria:

0-50% cell staining (negative)

51-100% cell staining (positive).

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result (Scale 0-4+).

The background intensity must be within 0.5 of the negative control slide (Scale 0-4+).

The intensity of the negative reagent control slide must be 0.5 or less (Scale 0-4+).

Bioburden	Pass	
-----------	------	--

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Guadalupe Villarreal *Guadalupe Villarreal GV*

Issue date

09 APR 2018