



Certificate of Analysis

Product name

VENTANA ANTI-P120 CATENIN (98)

Material No.

05867088001

Lot No.

N01960

Date of Manufacturing

2025-02-14

Expiry Date

2027-02-13

Parameter**Result****Unit**

Visual Inspection
Reagent must be clear of foreign material and particulate matter, and free of turbidity.

Pass

Functional Staining

Pass

Ductal and Lobular Breast Carcinoma Tissues

The test batch and reference batch must score 2.0 or greater on both ductal and lobular breast carcinoma tissues for specific stain intensity on a 0 to 3 scale.

The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch.

The background intensity of the test batch, reference batch, and negative reagent control must be 0.5 or less on both tissues

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Antonio Reyes

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2025-03-10

Page 1 of 1

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>