



Certificate of Analysis

Product name

VENTANA ANTI-P120 CATENIN (98)

Material No.

05867088001

Lot No.

Y08418

Date of Manufacturing

2017-05-22

Expiry Date

2019-05-22

Parameter**Result****Unit**

Functional Staining

Pass

Ductal and lobular breast carcinoma

The test batch must score 2.0 or greater on both ductal and lobular breast carcinoma tissue

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result

The background intensity must be within 0.5 of the negative control slide

The intensity of the negative reagent control slide must be 0.5 or less

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter

Issue date 2017-06-05

Page 1 of 2



Certificate of Analysis

Product name

VENTANA ANTI-P120 CATENIN (98)

Material No.

05867088001

Lot No.

Y08418

Date of Manufacturing

2017-05-22

Expiry Date

2019-05-22

Parameter**Result****Unit**

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Ken Eblen

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2017-06-05

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>