

CERTIFICATE OF ANALYSIS

Product Name: VENTANA anti-p63 (4A4) Mouse Monoclonal Primary Antibody
 Ventana REF: 790-4509
 Roche Number: 05867061001
 Batch Number: F03177

Test	Acceptance Criteria	Result
Functional Staining on normal tonsil uniblock and prostate carcinoma uniblock	Specific staining in normal cells of prostate carcinoma tissue must score 3.0 or greater on a scale of 0-4+	PASS
	Specific staining in neoplastic cells must score less than or equal to 0.5 on a scale of 0-4+	PASS
	Specific staining in the tonsil must score 3.0 or greater on a scale of 0-4+	PASS
	Test lot specific staining must be within 0.5 pt. of the reference lot	PASS
	The intensity of the negative control slide must be 0.5 or less	PASS
	Non-specific staining reading must score 0.5+ or less on a scale of 0-4+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2017/05/28

Manufacture date: 12/MAY/2015

Ken Eblen 

03-AUG-2015

Quality Department

Date