

**CERTIFICATE OF ANALYSIS**

---

Product: CONFIRM™ anti-p53 (DO7) Primary Antibody  
Ventana REF: 790-2912  
Roche Number: 05278074001  
Batch Number: E10278

Test	Acceptance Criteria	Result
Functional Staining on breast carcinoma or colon tissue	Reference slide reading must score 3.5+ or greater on a scale of 0-4+	PASS
	Background reading must score 0.5+ or less on a scale of 0-4+	PASS
	Intensity difference between sample and a reference must be within 0.5	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

---

This lot met the specifications.

Expiry date: 2017-01-28

Manufacture date : 26-JAN-2015

Guadalupe Villarreal

*Guadalupe Villarreal* 02 Feb 2015

**Quality Department**

**Date**