



cobas[®] 8000 modular analyzer series

Addendum: Specifications of radio equipment – Version 1.0
Software Version 06-03



cobas[®]
Life needs answers

Publication information

Publication version	Revision date	Change description
1.0	May-2017	First version

☰ Revision history

Edition notice

This addendum contains an addition to the **cobas**[®] 8000 Operator's Manual and is valid only for the **c** 701, **c** 702, and **e** 801 modules. This addendum adds specifications for radio equipment.

Copyright

© 2007–2017, Roche Diagnostics GmbH.
All rights reserved.

Trademarks

The following trademarks are acknowledged:

COBAS, COBAS C, COBAS E, COBAS INTEGRA, and LIFE NEEDS ANSWERS are trademarks of Roche.

All other trademarks are the property of their respective owners.

Approvals

The **cobas**[®] 8000 modular analyzer series meets the requirements laid down in:

Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices.

Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC.

The full text of the 2014/53/EU declaration of conformity is available at the following internet address:
<http://e-labdoc.roche.com>

Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

Compliance is provided by means of the Declaration of Conformity.

The following marks demonstrate compliance:



For *in vitro* diagnostic use.



Complies with the IVD directive 98/79/EC on *in vitro* diagnostic medical devices



Issued by Intertek for Canada and the US.

Contact addresses

Inside the European Union and EFTA member states



Manufacturer of the instrument

Hitachi High-Technologies Corporation
1-24-14 Nishi-Shimbashi
Minato-ku Tokyo 105-8717
Japan



Manufacturer of **cobas**[®] 8000 data manager

Roche Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim
Germany



Authorized representative

Roche Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim
Germany

Outside the European Union and EFTA member states

Manufactured by:

Hitachi High-Technologies Corporation

Manufactured for:

Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim
Germany

Table of contents

Revision 1: Specifications of radio equipment added	6
Specifications of radio equipment	6

Revision 1: Specifications of radio equipment added

In the chapter *Specifications*, the specifications of the RFID readers of the **c 701**, **c 702**, and **e 801** modules have been added.

Specifications of radio equipment

The following modules contain RFID readers:

	c 701 module	c 702 module	e 801 module
Frequency (MHz)	13.56	13.56	13.56
Maximum radio-frequency power (mW)	< 90	< 100	< 200
Number of RFID readers	2	1	8

☒ Specifications of radio equipment