



Certificate of Analysis

Product name

CONFIRM anti-Thyroid transcription fact

Material No.

05479312001

Lot No.

K13628

Date of Manufacturing

2023-07-20

Expiry Date

2025-07-09

Parameter	Result	Unit
Visual Inspection	Pass	
Reagent must be clear of foreign material and particulate matter, and free of turbidity.		
Functional Staining	Pass	
Normal Lung and Normal Thyroid Tissues		
For specific stain intensity, the test batch and reference batch must score 3.5 or greater on normal thyroid tissue and score 2 or greater on normal lung tissue on a 0 to 4 scale.		
The specific stain intensity of the test batch must also be within 0.5 pt. of the reference batch.		
The background intensity must be 0.5 or less on both tissues for the test batch and reference batch.		
The background intensity of the negative reagent control on both tissues must be 0.5 or less.		
Bioburden	Pass	
Visible bacterial colonies must be 100 or less CFU/mL		
This lot meets our specification		

Issue date 2023-08-08

Page 1 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>



Certificate of Analysis

Product name

CONFIRM anti-Thyroid transcription fact

Material No.

05479312001

Lot No.

K13628

Date of Manufacturing

2023-07-20

Expiry Date

2025-07-09

Disposition by

Mia Cazares

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2023-08-08

Page 2 of 2

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 800 227 2155
<http://www.ventana.com>