



a member of the Roche group

CERTIFICATE OF ANALYSIS

Product: PATHWAY[®] anti-HER-2/neu (4B5) Rabbit Monoclonal Primary Antibody
Ventana **REF**: 790-2991
Roche Number: 05278368001
Batch Number: F06373

Test	Acceptance Criteria	Result
Functional Staining on Cell line or Breast Cancer	Appropriate tissue specific staining on Levels 0-3	PASS
	Background reading must score 0.5+ or less on a scale of 0-3+	PASS
	Negative control staining must be negative for specific staining	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2017-03-28

Manufacture date : 08-Sep-2015

Ken Eblen  05-JAN-2016

Quality Department

Date