



Certificate of Analysis

Product name

CONFIRM Anti-PR (1E2) Primary Antibody

Material No.

05277990001 or 05278392001

Lot No.

E07608

Date of Manufacturing

16-Apr-2018

Expiry Date

2020/04/02

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be free of foreign material and substantially free of particulate matter

Reagent must be clear and free of turbidity

Functional Staining

Pass

Breast Carcinoma Tissue

The nuclear staining will be interpreted according to the following approximate percent tumor cell staining criteria:

0-50% cell staining (negative)

51-100% cell staining (positive).

The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result (Scale 0-4+).

The background intensity must be within 0.5 of the negative control slide (Scale 0-4+).

The intensity of the negative reagent control slide must be 0.5 or less (Scale 0-4+).

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

KEN ROCHIN

Issue date

02-MAY-2018