



## Certificate of Analysis

**Product name**

VENTANA PTEN (SP218)

**Material No.**

07970200001

**Lot No.**

E12810

**Date of Manufacturing**

2018-06-13

**Expiry Date**

2020-06-12

**Parameter****Result****Unit**

Visual Inspection

Pass

Reagent must be clear, and free of turbidity.

Reagent must be free of foreign material and substantially free of intrinsic particulate matter.

This lot meets our specification

**Disposition by**

Lupe Villarreal

This certificate is a computer printout and has therefore not been signed by hand.

**Issue date** 2018-06-29

Page 1 of 1

Ventana Medical Systems, Inc.  
1910 Innovation Park Drive  
Tucson, Arizona 85755  
Phone: + 1 520 887 2155  
<http://www.ventana.com>