



a member of the Roche group

CERTIFICATE OF ANALYSIS

Product Name: CONFIRM® anti-Topoisomerase II α (JS5B4) Rabbit Monoclonal Primary Antibody

Ventana [REF]: 790-4371

Roche Number: 05479339001

Batch Number: F05192

Test	Acceptance Criteria	Result
Functional Staining on enter tissue type	The test batch must score 3.5 or greater on breast cancer tissue	PASS
	The specific stain intensity must also be within 0.5 pt. of the reference batch for a passing result	PASS
	The background intensity must be within 0.5 of the negative control slide	PASS
	The intensity of the negative reagent control slide must be 0.5 or less	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiration date: 2017/07/28

Manufacture date: 31/Jul/2015

Lucia Romero Lucia Romero LR 19-AUG-2015
Quality Department **Date**