

**CERTIFICATE OF ANALYSIS**

---

Product: FITC anti-Lambda Primary Antibody  
Ventana **REF**: 760-2684  
Roche Number: 05267951001  
Batch Number: G02330

Test	Acceptance Criteria	Result
Functional Staining on tonsil or skin tissue	Tonsil or skin slide staining intensity must be equivalent or stronger to the slide stained with the reference lot	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal 100 CFU/mL	PASS

---

This lot met the specifications.

Expiry Date: 2018/03/28

Manufacture Date : 30/MAR/2016

Lucia Romero

  
Quality Department

 18-APR-2016  
Date