



Certificate of Analysis

Product name

PATHWAY anti-HER2/neu (4B5) Rabbit Monoclonal Primary Antibody

Material No.

05278368001

Lot No.

E17552

Date of Manufacturing

15-AUG-2018

Expiry Date

2020-01-23

Parameter**Result****Unit**

Visual Inspection

Pass

Reagent must be free of foreign material and substantially free of particulate matter

Reagent must be clear and free of turbidity

Functional Staining

Pass

Breast Carcinoma Tissue, Gastric Carcinoma Tissue, and 4 in 1 Cell Line.

The specific staining of the test batch must be acceptable on all tissues and cell line cores using both the *View* DAB IHC and the *ultraView* DAB IHC detection kits.

The test batch background intensity must not exceed 0.5 pt on any tissue or cell line core.

All tissues and cell line cores stained with the negative control reagent must be negative for specific staining

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

This lot meets our specification

Disposition by

Lucia Romero

Issue date

24-AUG-2018