

CERTIFICATE OF ANALYSIS

Product Name: HPV 16 mRNA Probe, ASR

HPV 18 mRNA Probe, ASR

HPV 33 mRNA Probe, ASR

HPV 6 mRNA Probe, ASR

HPV 11 mRNA Probe, ASR

Ventana [REF]: 760-1236, 760-1237, 760-1238, 760-1239, 760-1240

Roche Number: 07658834001, 07658842001, 07658869001, 07658877001, 07658885001

Batch Number: G03390Z

Test	Acceptance Criteria	Result
Functional Staining on sense RNA Dot Blot	The test batch must exhibit a positive spot for the target sequence	PASS
	All off target sequences must not show any spot. The exception being HPV 6 and HPV 11, which may show faint cross reactivity with each other, but no other genotypes.	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS

This lot met the specifications.

Expiration date: 2017/05/28

Manufacture date: 04-May-2016

Lucia Romero

Quality Department

Date