

**CERTIFICATE OF ANALYSIS**


---

Product Name: VENTANA anti-p120 catenin (98) Mouse Monoclonal Primary Antibody  
 Ventana REF: 790-4517  
 Roche Number: 05867088001  
 Batch Number: E01726

Test	Acceptance Criteria	Result
Functional Staining on p120-positive ductal and lobular carcinoma tissues containing normal and neoplastic regions	The specific staining of the test lot must score 2.0 or greater in both the ductal and lobular carcinoma (scale 0-3+)	PASS
	The test lot specific stain intensity on both the ductal and lobular carcinoma must be within 0.5 pt. of the reference lot	PASS
	The intensity of the negative reagent control slide must be 0.5 or less	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2016-03-28

Manufacture date: 27-Mar-2014

Laughlin Hoskins:

*Laughlin Hoskins 02-Apr-2014*

**Quality Department**

**Date**