


Mat-Nr.: 04687787190
HBsAg G2 Elecsys cobas e
Batch: 11054903
2° 8°C
Lot: 040000032659 08.02.2016
Sample No./Type: 200034688/Y2



Verification Report

according to Directive 98/79/EC Annex IV.6

No. ROC-04687787 16 03 034959

Manufacturer: ROCHE Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim

Product: HBsAg II
Elecsys and cobas e analyzers

Batch: 11054903

Expiry Date: 08-2016

The above mentioned batch meets the batch release criteria established during design examination and may be placed on the market. The design examination certificate issued for this product is V7 13 02 10283 285.

Date, 2016-03-02

i.A. Dr. Volker Schlüter
In-vitro-Diagnostika

TÜV SÜD Product Service GmbH is Notified Body according to Council Directive 98/79/EC concerning In Vitro Diagnostic Medical Devices with Identification No. 0123.

To (via Telefax):

ROCHE Diagnostics GmbH
Dr. Rainer Bäuerlein
08856 / 60-3647

CC: ---

Headquarters: Munich
Trade Register Munich
HRB 85742
V.A.T. DE 129484287
UniCredit Bank AG
BIC HYVEDEMMXXX
IBAN DE13 7002 0270 0048 8522 11

Managing Directors:
Robert Keas (CEO)
Dr. Jens Butenandt

Tel.: +49 89 5008-4483
Fax: +49 89 5008-4475
www.tuev-sued.de
TÜV®

TÜV SÜD Product Service GmbH
In-vitro-Diagnostika
Ridlerstraße 65
80339 München
Germany