



Certificate of Analysis

Product name

PROTEASE 3

Material No.

05266718001

Lot No.

E22160

Date of Manufacturing

2018-09-27

Expiry Date

2020-03-04

Parameter**Result****Unit**

Absorbance

Pass

Reference and Sample Mean must score in the range of 0.010 and 0.055

Bioburden

Pass

Visible bacterial colonies must be 100 or less CFU/mL

Visual Inspection

Pass

Product should be clear and colorless, but it is acceptable for the reagent to contain a white flocculent material of a non-biological origin.

This lot meets our specification

Disposition by

Lucia Romero

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2018-10-12

Page 1 of 1

Ventana Medical Systems, Inc.
1910 Innovation Park Drive
Tucson, Arizona 85755
Phone: + 1 520 887 2155
<http://www.ventana.com>