



a member of the Roche group

CERTIFICATE OF ANALYSIS

Product Name: VENTANA anti-p63 (4A4) Mouse Monoclonal Primary Antibody
Ventana REF: 790-4509
Roche Number: 05867061001
Batch Number: D04852

| Test | Acceptance Criteria | Result |
|---|---|--------|
| Functional Staining on normal tonsil uniblock and prostate carcinoma uniblock | Specific staining in normal cells of prostate carcinoma tissue must score 3.0 or greater on a scale of 0-4+ | PASS |
| | Specific staining in neoplastic cells must score less than or equal to 0.5 on a scale of 0-4+ | PASS |
| | Specific staining in the tonsil must score 3.0 or greater on a scale of 0-4+ | PASS |
| | Test lot specific staining must be within 0.5 pt. of the reference lot | PASS |
| | The intensity of the negative control slide must be 0.5 or less | PASS |
| | Non-specific staining reading must score 0.5+ or less on a scale of 0-4+ | PASS |
| Visual Inspection | Reagent must be free of foreign material and substantially free of intrinsic particulate matter | PASS |
| | Reagent must be clear and free of turbidity | PASS |
| Bioburden | Visible bacterial colonies must be less than or equal to 100 CFU/mL | PASS |

This lot met the specifications.

Expiry date: 2014-06-28

Manufacture date: 28-Jun-2013

Angela Glenn
Quality Department

27-Jun-2013
Date