

CERTIFICATE OF ANALYSIS

Product Name: VENTANA anti-E-cadherin (36) Mouse Monoclonal Primary Antibody
 Ventana REF: 790-4497
 Roche Number: 05905290001
 Batch Number: E01403

Test	Acceptance Criteria	Result
Functional Staining ductal breast and lung carcinoma	Reference slide reading must score 2.5+ or greater on a scale of 0-3+	PASS
	Background reading must score 0.5 or less on a scale of 0-4+	PASS
	The specific stain intensity must be within 0.5 pt. of the reference batch for a passing result.	PASS
	Intensity of the negative reagent control slide must be 0.5 or less on a scale of 0-4+	PASS
Visual Inspection	Reagent must be free of foreign material and substantially free of intrinsic particulate matter	PASS
	Reagent must be clear and free of turbidity	PASS
Bioburden	Visible bacterial colonies must be less than or equal to 100 CFU/mL	PASS

This lot met the specifications.

Expiry date: 2016-03-28

Manufacture date: 19-MAR-2014

Guadalupe Villarreal

Guadalupe Villarreal

24 MAR 2014

Quality Department

Date