



CERTIFICATE OF ANALYSIS

Product Name: HPV 16 mRNA Probe, ASR

HPV 18 mRNA Probe, ASR HPV 33 mRNA Probe, ASR HPV 6 mRNA Probe, ASR

HPV 11 mRNA Probe, ASR

Ventana REF: 760-1236, 760-1237, 760-1238, 760-1239, 760-1240

Roche Number: 07658834001, 07658842001, 07658869001, 07658877001, 07658885001

Batch Number: G03377Z

| Test | Acceptance Criteria | Result |
|---|---|--------|
| Functional Staining on sense RNA Dot Blot | The test batch must exhibit a positive spot for the target sequence | PASS |
| | All off target sequences must not show any spot. The exception being HPV 6 and HPV 11, which may show faint cross reactivity with each other, but no other genotypes. | PASS |
| Visual Inspection | Reagent must be free of foreign material and substantially free of particulate matter | PASS |
| | Reagent must be clear and free of turbidity | PASS |
| | | |

This lot met the specifications.

Expiration date: 2017/04/28

Manufacture date: 26-APR-2016

aughlin Horrine 25-may-2016

Quality Department